

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 30, 2001
LDC Report Date: August 24, 2001
Matrix: Water
Parameters: Volatiles
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory
Sample Delivery Group (SDG): 01-5003

Sample Identification

ER-14
MW-14-1
MW-14-5
MW-14-4
MW-14-3
MW-14-2
MW-14-2D
TRIP BLANK
MW-14-3MS
MW-14-3MSD

Introduction

This data review covers 10 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 524.2 for Volatiles.

This review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Organic Data Review (October 1999) as there are no current guidelines for the method stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section V.

Field duplicates are summarized in Section XVI.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. GC/MS Instrument Performance Check

Instrument performance was checked at 12 hour intervals.

All ion abundance requirements were met.

III. Initial Calibration

Initial calibration was performed using required standard concentrations.

Percent relative standard deviations (%RSD) were less than or equal to 20.0% for selected compounds.

A curve fit, based on the initial calibration, was established for quantitation for selected compounds. The coefficient of determination (r^2) was greater than or equal to 0.990 with the following exceptions:

Date	Compound	r^2	Associated Samples	Flag	A or P
7/19/01	Chloromethane	0.9897	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 TRIP BLANK MW-14-3MS MW-14-3MSD 01G3782MB01	J (all detects) UJ (all non-detects)	P

IV. Continuing Calibration

Continuing calibration was performed at the required frequencies.

All of the continuing calibration percent differences (%D) between the initial calibration RRF and the continuing calibration RRF were less than or equal to 30.0% with the following exceptions:

Date	Compound	%D	Associated Samples	Flag	A or P
8/1/01	1,2,3-Trichloropropane	32.66	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 TRIP BLANK MW-14-3MS MW-14-3MSD 01G3782MB01	J (all detects) UJ (all non-detects)	P

V. Blanks

Method blanks were reviewed for each matrix as applicable. No volatile contaminants were found in the method blanks with the following exceptions:

Method Blank ID	Analysis Date	Compound TIC (RT in minutes)	Concentration	Associated Samples
01G3782MB01	8/2/01	Methylene chloride	5.2 ug/L	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 TRIP BLANK
01G3890MB01	8/13/01	Methylene chloride	1.6 ug/L	MW-14-2D

Sample concentrations were compared to concentrations detected in the method blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated method blanks with the following exceptions:

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
MW-14-1	Methylene chloride	1 ug/L	1U ug/L
MW-14-4	Methylene chloride	1.3 ug/L	1.3U ug/L
MW-14-3	Methylene chloride	1.8 ug/L	1.8U ug/L
MW-14-2	Methylene chloride	1.7 ug/L	1.7U ug/L
TRIP BLANK	Methylene chloride	2.3 ug/L	2.3U ug/L

Sample	Compound TIC (RT in minutes)	Reported Concentration	Modified Final Concentration
MW-14-2D	Methylene chloride	2.1 ug/L	2.1U ug/L

Sample ER-14 was identified as an equipment rinsate. No volatile contaminants were found in this blank with the following exceptions:

Equipment Rinsate ID	Sampling Date	Compound	Concentration	Associated Samples
ER-14	7/30/01	Chloroform Bromodichloromethane	0.3 ug/L 0.4 ug/L	MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 MW-14-2D

Sample "TRIP BLANK" was identified as a trip blank. No volatile contaminants were found in this blank with the following exceptions:

Trip Blank ID	Sampling Date	Compound	Concentration	Associated Samples
TRIP BLANK	7/30/01	Methylene chloride	2.3 ug/L	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 MW-14-2D

Sample concentrations were compared to concentrations detected in the field blanks. The sample concentrations were either not detected or were significantly greater (>10X for common contaminants, >5X for other contaminants) than the concentrations found in the associated field blanks with the following exceptions:

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-14-3	Chloroform Methylene chloride	0.6 ug/L 1.8 ug/L	0.6U ug/L 1.8U ug/L
MW-14-2	Chloroform Methylene chloride	0.6 ug/L 1.7 ug/L	0.6U ug/L 1.7U ug/L
MW-14-2D	Chloroform Methylene chloride	0.6 ug/L 2.1 ug/L	0.6U ug/L 2.1U ug/L

Sample	Compound	Reported Concentration	Modified Final Concentration
MW-14-1	Methylene chloride	1 ug/L	1U ug/L
MW-14-4	Methylene chloride	1.3 ug/L	1.3U ug/L

VI. Surrogate Spikes

Surrogates were added to all samples and blanks as required by the method. All surrogate recoveries (%R) were within QC limits.

VII. Matrix Spike/Matrix Spike Duplicates

Although matrix spike (MS) and matrix spike duplicate (MSD) samples were not required by the method, MS and MSD samples were reported by the laboratory. Percent recoveries (%R) and relative percent differences (RPD) were within QC limit.

VIII. Laboratory Control Samples (LCS)

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

IX. Regional Quality Assurance and Quality Control

Not applicable.

X. Internal Standards

All internal standard areas and retention times were within QC limits.

XI. Target Compound Identifications

All target compound identifications were within validation criteria.

XII. Compound Quantitation and CRQLs

All compound quantitation and CRQLs were within validation criteria.

XIII. Tentatively Identified Compounds (TICs)

Tentatively identified compounds were not reported by the laboratory.

XIV. System Performance

The system performance was acceptable.

XV. Overall Assessment of Data

Data flags have been summarized at the end of the report.

XVI. Field Duplicates

Samples MW-14-2 and MW-14-2D were identified as field duplicates. No volatiles were detected in any of the samples with the following exceptions:

Compound	Concentration (ug/L)		RPD
	MW-14-2	MW-14-2D	
Chloroform	0.6	0.6	0
1,1-Dichloroethane	0.4	0.5	22
Methylene chloride	1.7	2.1	21
Tetrachloroethene	0.8	0.8	0
1,2,3-Trichlorobenzene	1.2	0.9	29
Trichloroethene	2.7	3.4	23

JPL, 00HW019

Volatiles - Data Qualification Summary - SDG 01-5003

SDG	Sample	Compound	Flag	A or P	Reason
01-5003	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 TRIP BLANK	Chloromethane	J (all detects) UJ (all non-detects)	P	Initial calibration (r ²)
01-5003	ER-14 MW-14-1 MW-14-5 MW-14-4 MW-14-3 MW-14-2 TRIP BLANK	1,2,3-Trichloropropane	J (all detects) UJ (all non-detects)	P	Continuing calibration (%D)

JPL, 00HW019

Volatiles - Laboratory Blank Data Qualification Summary - SDG 01-5003

SDG	Sample	Compound TIC (RT in minutes)	Modified Final Concentration	A or P
01-5003	MW-14-1	Methylene chloride	1U ug/L	A
01-5003	MW-14-4	Methylene chloride	1.3U ug/L	A
01-5003	MW-14-3	Methylene chloride	1.8U ug/L	A
01-5003	MW-14-2	Methylene chloride	1.7U ug/L	A
01-5003	TRIP BLANK	Methylene chloride	2.3U ug/L	A
01-5003	MW-14-2D	Methylene chloride	2.1U ug/L	A

JPL, 00HW019

Volatiles - Field Blank Data Qualification Summary - SDG 01-5003

SDG	Sample	Compound	Modified Final Concentration	A or P
01-5003	MW-14-3	Chloroform Methylene chloride	0.6U ug/L 1.8U ug/L	A
01-5003	MW-14-2	Chloroform Methylene chloride	0.6U ug/L 1.7U ug/L	A
01-5003	MW-14-2D	Chloroform Methylene chloride	0.6U ug/L 2.1U ug/L	A
01-5003	MW-14-1	Methylene chloride	1U ug/L	A
01-5003	MW-14-4	Methylene chloride	1.3U ug/L	A

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 6, 2001
LDC Report Date: August 14, 2001
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4565

Sample Identification

ER-17
MW-17-2
MW-17-3
MW-17-4
MW-17-5

Introduction

This data review covers 5 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4565	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

No field blanks were identified in this SDG.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Wet Chemistry - Data Qualification Summary - SDG 01-4565

SDG	Sample	Analyte	Flag	A or P	Reason
01-4565	ER-17 MW-17-2 MW-17-3 MW-17-4 MW-17-5	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 01-4565

No Sample Data Qualified in this SDG

JPL, 00HW019
Wet Chemistry - Field Blank Data Qualification Summary - SDG 01-4565

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 9, 2001
LDC Report Date: August 15, 2001
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4599

Sample Identification

ER-21
MW-21-1
MW-21-2
MW-21-3
MW-21-4
MW-21-5

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UU Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4599	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-21 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Perchlorate - Data Qualification Summary - SDG 01-4599

SDG	Sample	Analyte	Flag	A or P	Reason
01-4599	ER-21 MW-21-1 MW-21-2 MW-21-3 MW-21-4 MW-21-5	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Perchlorate - Laboratory Blank Data Qualification Summary - SDG 01-4599

No Sample Data Qualified in this SDG

JPL, 00HW019
Perchlorate - Field Blank Data Qualification Summary - SDG 01-4599

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 10 through July 11, 2001
LDC Report Date: August 14, 2001
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4650

Sample Identification

MW-6
MW-10
MW-13
MW-10-D
MW-20-5
MW-6MS
MW-6MSD
MW-13MS
MW-13MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met with the following exceptions:

Sample	Analyte	Total Time From Sample Collection Until Analysis	Required Holding Time From Sample Collection Until Analysis	Flag	A or P
MW-20-5	Hexavalent chromium	44.75 hrs.	24 hrs.	J (all detects) UJ (all non-detects)	P

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4650	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

No field blanks were identified in this SDG.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

Samples MW-10 and MW-10-D were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	MW-10	MW-10-D	
Perchlorate	24	30	22

JPL, 00HW019
Wet Chemistry - Data Qualification Summary - SDG 01-4650

SDG	Sample	Analyte	Flag	A or P	Reason
01-4650	MW-20-5	Hexavalent chromium	J (all detects) UJ (all non-detects)	P	Technical holding times
01-4650	MW-6 MW-10 MW-13 MW-10-D MW-20-5	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 01-4650

No Sample Data Qualified in this SDG

JPL, 00HW019
Wet Chemistry - Field Blank Data Qualification Summary - SDG 01-4650

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 12, 2001
LDC Report Date: August 15, 2001
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4684

Sample Identification

MW-5
MW-16
MW-5-D
MW-16-D
MW-5MS
MW-5MSD

Introduction

This data review covers 6 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4684	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

No field blanks were identified in this SDG.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

Samples MW-16 and MW-16-D, and samples MW-5 and MW-5D were identified as field duplicates. No contaminant concentrations were detected in any of the samples with the following exceptions:

Analyte	Concentration (ug/L)		RPD
	MW-16	MW-16-D	
Perchlorate	1800	1800	0

JPL, 00HW019
Wet Chemistry - Data Qualification Summary - SDG 01-4684

SDG	Sample	Analyte	Flag	A or P	Reason
01-4684	MW-5 MW-16 MW-5-D MW-16-D	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 01-4684

No Sample Data Qualified in this SDG

JPL, 00HW019
Wet Chemistry - Field Blank Data Qualification Summary - SDG 01-4684

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 13, 2001
LDC Report Date: August 15, 2001
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4685

Sample Identification

ER-3
MW-3-2
MW-3-3
MW-3-4
MW-3-5
MW-3-5MS
MW-3-5MSD

Introduction

This data review covers 7 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4685	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-3 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Wet Chemistry - Data Qualification Summary - SDG 01-4685

SDG	Sample	Analyte	Flag	A or P	Reason
01-4685	ER-3 MW-3-2 MW-3-3 MW-3-4 MW-3-5	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 01-4685

No Sample Data Qualified in this SDG

JPL, 00HW019
Wet Chemistry - Field Blank Data Qualification Summary - SDG 01-4685

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 16, 2001
LDC Report Date: August 15, 2001
Matrix: Water
Parameters: Wet Chemistry
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4729

Sample Identification

ER-20
MW-20-1
MW-20-2
MW-20-3
MW-20-4
ER-20MS
ER-20MSD
MW-20-4MS
MW-20-4MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate and EPA SW 846 Method 7196 for Hexavalent Chromium.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4729	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-20 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

No field duplicates were identified in this SDG.

JPL, 00HW019
Wet Chemistry - Data Qualification Summary - SDG 01-4729

SDG	Sample	Analyte	Flag	A or P	Reason
01-4729	ER-20 MW-20-1 MW-20-2 MW-20-3 MW-20-4	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Wet Chemistry - Laboratory Blank Data Qualification Summary - SDG 01-4729

No Sample Data Qualified in this SDG

JPL, 00HW019
Wet Chemistry - Field Blank Data Qualification Summary - SDG 01-4729

No Sample Data Qualified in this SDG

**Laboratory Data Consultants, Inc.
Data Validation Report**

Project/Site Name: JPL, 00HW019
Collection Date: July 18, 2001
LDC Report Date: August 15, 2001
Matrix: Water
Parameters: Perchlorate
Validation Level: EPA Level IV
Laboratory: Applied P & Ch Laboratory

Sample Delivery Group (SDG): 01-4770

Sample Identification

ER-19
MW-19-5
MW-19-4
MW-19-1
MW-19-3
MW-19-2
MW-19-2D
MW-19-3MS
MW-19-3MSD

Introduction

This data review covers 9 water samples listed on the cover sheet including dilutions and reanalysis as applicable. The analyses were per EPA Method 314.0 for Perchlorate.

The review follows a modified outline of the USEPA Contract Laboratory Program National Functional Guidelines for Inorganic Data Review (February 1994) as there are no current guidelines for the methods stated above.

A table summarizing all data qualification is provided at the end of this report. Flags are classified as P (protocol) or A (advisory) to indicate whether the flag is due to a laboratory deviation from a specified protocol or is of technical advisory nature.

Blank results are summarized in Section III.

Field duplicates are summarized in Section VII.

The following are definitions of the data qualifiers:

- U Indicates the compound or analyte was analyzed for but not detected at or above the stated limit.
- J Indicates an estimated value.
- R Quality control indicates the data is not usable.
- N Presumptive evidence of presence of the constituent.
- UJ Indicates the compound or analyte was analyzed for but not detected. The sample detection limit is an estimated value.
- A Indicates the finding is based upon technical validation criteria.
- P Indicates the finding is related to a protocol/contractual deviation.
- None Indicates the data was not significantly impacted by the finding, therefore qualification was not required.

I. Technical Holding Times

All technical holding time requirements were met.

The chain-of-custodies were reviewed for documentation of cooler temperatures. All cooler temperatures met validation criteria.

II. Calibration

a. Initial Calibration

All criteria for the initial calibration of each method were met with the following exceptions:

Sample	Analyte	Finding	Criteria	Flag	A or P
All samples in SDG 01-4770	Perchlorate	A blank was not used to establish the calibration curve.	A blank must be used to establish the calibration curve.	None	P

b. Calibration Verification

Calibration verification frequency and analysis criteria were met for each method when applicable.

III. Blanks

Method blanks were reviewed for each matrix as applicable. No contaminant concentrations were found in the method blanks.

Sample ER-19 was identified as an equipment rinsate. No contaminant concentrations were found in this blank.

IV. Accuracy and Precision Data

a. Matrix Spike/(Matrix Spike) Duplicates

Matrix spike (MS) and matrix spike duplicate (MSD) analyses were reviewed for each matrix as applicable. Percent recoveries (%R) and relative percent differences (RPD) were within QC limits.

b. Laboratory Control Samples

Laboratory control samples were reviewed for each matrix as applicable. Percent recoveries (%R) were within QC limits.

V. Sample Result Verification

All sample result verifications were within validation criteria.

VI. Overall Assessment of Data

Data flags are summarized at the end of this report.

VII. Field Duplicates

Samples MW-19-2 and MW-19-2D were identified as field duplicates. No contaminant concentrations were detected in any of the samples.

JPL, 00HW019
Perchlorate - Data Qualification Summary - SDG 01-4770

SDG	Sample	Analyte	Flag	A or P	Reason
01-4770	ER-19 MW-19-5 MW-19-4 MW-19-1 MW-19-3 MW-19-2 MW-19-2D	Perchlorate	None	P	Initial calibration

JPL, 00HW019
Perchlorate - Laboratory Blank Data Qualification Summary - SDG 01-4770

No Sample Data Qualified in this SDG

JPL, 00HW019
Perchlorate - Field Blank Data Qualification Summary - SDG 01-4770

No Sample Data Qualified in this SDG